



V.14/Aug/2016

# Egg (Ovalbumin) ELISA Kit

For The Quantitative Determination for  
Egg Protein in Food

**Catalog #M2101  
96 Assays**

***For Research or Laboratory Use Only.  
Not for Use in Diagnostic Procedures.  
Please read full descriptions in this manual before use.***

E-mail: [info@miobs.com](mailto:info@miobs.com)  
Website: <http://www.miobs-e.com/index.html>

JP.PAT.P3600231

## TABLE OF CONTENTS

INTENDED USE AND USER.....	3
PRINCIPLE OF ELISA .....	3
KIT COMPONENTS .....	3
REQUIRED MATERIALS (NOT INCLUDED) .....	4
PERFORMANCE CHARACTERISTICS .....	4
REAGENT PREPARATION .....	5
ASSAY PROTOCOL .....	6
<i>Summary of Procedure</i> .....	8
RESULTS AND INTERPRETATION .....	9
TYPICAL STANDARD CURVE .....	9
STORAGE CONDITIONS AND EXPIRY OF KIT .....	10
WARNINGS AND PRECAUTIONS .....	10
<i>Reagent preparation</i> .....	11
TECHNICAL ASSISTANCE .....	11
WARRANTY.....	11

For optimal results, follow instructions exactly. Failure to do so may lead to inaccurate results.

Reproducible results depend on careful pipetting technique, maintaining incubations at the specified temperature for the specified time, complete washing, and thorough mixing of all solutions.

### **INTENDED USE AND USER**

Egg (Ovalbumin) ELISA Kit is a sandwich enzyme immunoassay for the quantitative determination of egg protein in processed or unprocessed food. This kit is designed to be used by quality control personnel or other trained professionals.

### **PRINCIPLE OF ELISA**

Egg protein in the samples is extracted, centrifuged and filtered prior to the sandwich ELISA. The egg protein (antigen) is bound to the anti-egg protein polyclonal antibody coated wells of the microplate module. This results in the formation of an antigen-antibody complex in the wells. Unbound materials are removed by washing. Subsequently, the enzyme-conjugated anti-egg protein antibody is bound to the already bound egg protein-antibody complex, forming an antibody-antigen-antibody sandwich. The second washing step removes the excess conjugated antibody. Addition of enzyme substrate results in color development due to the enzyme bound to the complex. After addition of the stop solution, the color intensity of the solutions can be determined by the absorbance at 450 nm. The intensity of the color developed is directly proportional to the concentration of the egg protein in the food. The concentration of egg protein corresponding to the measured absorbance is determined by preparing a standard curve, and adjusting for a further dilution factor if necessary.

### **KIT COMPONENTS**

<b>Label</b>	<b>Name of component</b>	<b>Content</b>	<b>Quantity</b>
A	Antibody-coated Microplate Module	6X8 well modules	2 packs
B	Egg Standard (50ng/mL)	1 mL	2 vials
C	Enzyme-conjugated Antibody	13 mL	1 bottle
D	Enzyme Substrate (TMB Solution)	13 mL	1 bottle
E	Stop Solution (1N Sulfuric Acid)	13 mL	1 bottle
F	Sample Buffer (20X Concentrate)	50 mL	1 bottle
G	Wash Buffer (20X Concentrate)	50 mL	1 bottle
H	Extraction Component A (20X Concentrate)	55 mL	1 bottle
	Frame for mounting the microplate module		1 piece
	Microplate cover		1 piece

## **REQUIRED MATERIALS (NOT INCLUDED)**

1. 2-Mercaptoethanol
2. Distilled water (or deionized water)
3. Micropipettes and disposable tips ranging from 50 to 1000  $\mu$ L
4. Graduated cylinders
5. Polypropylene centrifuge tubes (50 mL size) for sample extraction
6. Polypropylene micro tubes (1–2 mL size) for preparation of working standard solutions and dilution of sample
7. Tube racks
8. pH test paper
9. Clear-plastic wrap
10. Homogenizer/blender for sample preparation (if necessary)
11. Water bath for boiling or horizontal shaker, for sample extraction
12. Centrifuge
13. Vortex mixer
14. Aspirator for washing procedure, or optional microplate washer
15. Microplate reader with a 450 nm filter, and a filter for any wavelength from 600 to 650 nm, inclusive

## **PERFORMANCE CHARACTERISTICS**

Egg (Ovalbumin) ELISA kit is a tool to determine the presence of egg protein in foods such as breakfast cereal, cookie, ice cream and milk chocolate.

Sample preparation time	Approx. 10 min	
Sample extraction time	Approx. 30 min, or overnight	
Time for ELISA analysis	Approx. 2.5 h	
Assay sensitivity	0.3 $\mu$ g egg protein/g food	
Detectable concentration range	0.78 to 50 ng/mL egg protein	
Intra-assay precision	< 10 %	
Inter-assay precision	< 10 %	
Recovery	100 +/- 25 %	
Specificity	Ingredient	Reactivity (%)
	Chicken raw egg	100
	Chicken boiled egg	90.3
	Chicken raw meat	Not detected
	Chicken boiled meat	Not detected

Additional specificity data are available. In need, please contact to Morinaga Institute of Biological Science, Inc. or your local distributor. Specificity results are the data obtained by commercially available materials, and vary somehow from lot to lot.

## **REAGENT PREPARATION**

### **Sample Extraction Solution**

Prepare the *Sample Extraction Solution* by mixing *Sample Buffer (F)*, *Extraction Component A (H)*, 2-Mercaptoethanol and distilled water at a ratio of **5:5:2:88**.

The following example would prepare enough *Sample Extraction Solution* to assay 40 samples:

<i>Sample Buffer (20X Concentrate) (F)</i>	40 mL
<i>Extraction Component A (20X Concentrate) (H)</i>	40 mL
2-Mercaptoethanol	16 mL
Distilled water	704 mL
<hr/>	
Total	800 mL

### **Note:**

a. *Extraction Component A (H)* may produce crystals after refrigerated storage. These crystals must be re-dissolved completely in a water bath at 30-37°C (86-99°F) prior to use. The fully re-dissolved *Extraction Component A* can be stored at 20-25°C (68-77°F).

b. *Sample Extraction Solution* can be stored at 4°C (39°F) for preparing *Diluent II*. If *Sample Extraction Solution* forms a precipitate after refrigerated storage, then the solution must be warmed in a water bath at 20-25°C (68-77°F) to re-dissolve the precipitate prior to use.

### **Diluent I**

*Diluent I* is used for preparing *Diluent II* and for diluting the *Sample Extract*.

Dilute *Sample Buffer (F)* 20-fold with distilled water.

The following is an example of preparing *Diluent I*:

<i>Sample Buffer (20X Concentrate) (F)</i>	5 mL
Distilled water	95 mL
<hr/>	
Total	100 mL

### **Diluent II**

*Diluent II* is used to prepare *Working Egg Standard*, and to further dilute the *Working Sample Solution*.

Dilute *Sample Extraction Solution*, 20-fold with *Diluent I*.

The following is an example of preparing *Diluent II*:

<i>Sample Extraction Solution</i>	2 mL
<i>Diluent I</i>	38 mL
<hr/>	
Total	40 mL

### **Washing Solution**

Dilute *Wash Buffer (G)* 20-fold with distilled water, and gently mix until the solution is homogeneous.

### **Egg Standard (50 ng/mL) (B)**

*Egg Standard (50 ng/mL) (B)* is used as a 50 ng/mL standard as well as for preparing the diluted *Working Egg Standard* described in ASSAY PROTOCOL b.

*Egg Standard* fills the criteria of egg reference material in Japan.

## **ASSAY PROTOCOL**

*It is strongly recommended to use disposable polypropylene tubes, and care should be taken to clean up all equipments and materials so as to be free from cross-contamination, since the assay is highly sensitive.*

### **a. Sample Preparation**

Extraction of egg protein from the test food sample can be performed by two different methods depending on customer's convenience.

#### ***(Short Time Extraction Method)***

1. Grind and mix up the test food sample to homogeneity with a contamination-free homogenizer/blender.
2. A disposable polypropylene centrifuge tube, containing 1.0 g of the homogenized sample and 19 mL of ***Sample Extraction Solution***, is capped tightly and vortexed for 30 seconds.
3. Expose the capped tube to 100°C (212°F) in boiling water for 10 minutes.
4. Place the tube in running water to cool it down to ambient temperature (it takes some 10 minutes).
5. Vortex the tube for 30 seconds.
6. Check the fluid pH with pH test paper, and neutralize (pH 6-8) with HCl or NaOH, if required.
7. Centrifuge the tube at 3,000 × g for 20 minutes at 20-25°C (68-77°F), and the supernatant is retained as ***Sample Extract***. (Filter the supernatant with filter paper, if necessary.)
8. Dilute the ***Sample Extract*** by 20-fold with ***Diluent I*** (see REAGENT PREPARATION), and the diluted solution is used as the ***Working Sample Solution*** for ELISA.

*Note: If further sample dilution is required, dilute the ***Working Sample Solution*** with **Diluent II**.*

#### ***(Overnight Extraction Method)***

1. Grind and mix up the test food sample to homogeneity with a contamination-free homogenizer/blender.
2. A disposable polypropylene centrifuge tube, containing 1.0 g of the homogenized sample and 19 mL of ***Sample Extraction Solution***, is capped tightly and vortexed for 30 seconds.
3. Fix the capped centrifuge tube to a shaker horizontally, and oscillate at room temperature overnight (for at least 12 hours at 90-110 rpm) with a reciprocating motion of about 3 cm.
4. Follow the *procedures 6-8* for ***Short Time Extraction Method***.

### **b. Working Egg Standard Preparation**

1. Dispense 0.5 mL ***Diluent II*** into six polypropylene micro tubes labeled 0.78, 1.56, 3.13, 6.25, 12.5 and 25 ng/mL.
2. Dispense 0.5 mL of 50 ng/mL ***Egg Standard (B)*** into the 25 ng/mL tube, and mix thoroughly.
3. Dispense 0.5 mL of the freshly-prepared 25 ng/mL standard into the 12.5 ng/mL tube, and mix thoroughly.
4. Repeat this 2-fold dilution series to prepare 6.25, 3.13, 1.56, and 0.78 ng/mL standards.
5. Dispense 0.5 mL of ***Diluent II*** into the polypropylene micro tube labeled 0 ng/mL.

### c. ELISA

#### **(First reaction)**

1. Unseal the **Antibody-coated Microplate Module (A)** after equilibrated to room temperature (20-25°C, 68-77°F). Fix the module to the mounting frame (supplied).
2. Into each well, dispense 100 µL of standards (0, 0.78, 1.56, 3.13, 6.25, 12.5, 25 and 50 ng/mL **Working Egg Standard**) and **Working Sample Solution**, in (at least) duplicate.  
*Note: Samples presumed to contain greater than 50 ng/mL of egg protein should be diluted with **Diluent II** (ASSAY PROTOCOL a.8, Note).*
3. Cover the microplate with a microplate cover (supplied), and wrap them all by clear-plastic wrap to minimize the odor during incubation.
4. Incubate for 1 hour at 20-25°C (68-77°F).

#### **(Second reaction)**

1. Carefully unwrap the plate, and then remove the solution completely from each well by aspiration. (OR: Carefully unwrap the plate, and pour the liquid out of the wells and tap the microplate frame upside down vigorously for five times on paper towels to ensure complete removal of liquid from wells.)
2. Wash six times, by adding 300 µL **Washing Solution** per well followed by aspiration. After washing, remove any remaining solution by inverting and tapping the plate on a clean paper towel. (OR: Fill all the wells with 300 µL of the **Washing Solution** and pour out the liquid again and tap. Repeat five more times.)
3. Add 100 µL **Enzyme-conjugated Antibody (C)** to each well.
4. Cover the microplate with a microplate cover and incubate for precisely 30 minutes at 20-25°C (68-77°F).

#### **(Enzyme reaction)**

1. Completely aspirate the well contents and wash six times by filling with 300 µL **Washing Solution** per well, and aspirate. After the sixth washing, remove any remaining solution by inverting and tapping the plate on a clean paper towel. (OR: Completely pour the liquid out of the wells, then fill all the wells with 300 µL of the **Washing Solution**, pour out the liquid again and tap. Repeat five more times.)
2. Immediately, dispense 100 µL **Enzyme Substrate (D)**.
3. Cover the microplate with the microplate cover and incubate for precisely 30 minutes at 20-25°C (68-77°F). During the enzyme reaction, avoid exposing the microplate to light.
4. Stop the enzyme reaction by adding 100 µL of **Stop Solution (E)**.
5. Immediately measure absorbance using a microplate reader (measure at 450 nm, subtract a reference wavelength between 600 and 650 nm, inclusive).

*Note: The absorbance must be measured within 30 minutes after stopping the enzyme reaction.*

## **Summary of Procedure**

### **(Extraction procedure)**

Grind/mince the samples to homogeneity

Weigh 1.0 g ground (or liquid) sample in a clean centrifuge tube, add 19 mL **Sample Extraction Solution**, cap the tube tightly, and vortex.

#### **Short Time Extraction Method**

- Expose the capped tube to 100°C (212°F) in boiling water for 10 minutes.
- Cool down the centrifuge tube in running water, and vortex for 30 sec.
- Adjust pH to 6.0-8.0.

#### **Overnight Extraction Method**

- Fix the capped centrifuge tube to a shaker horizontally, and oscillate at room temperature overnight (for at least 12 hours at 90-110 rpm) with a reciprocating motion of about 3 cm.
- Adjust pH to 6.0-8.0.

Centrifuge at 3,000 × g for 20 min at 20-25°C (68-77°F) and filter the supernatant if necessary.

Dilute **Sample Extract** 20-fold with **Diluent I**  
(If necessary, dilute further using **Diluent II**)

### **(ELISA procedure)**

Attach the **Antibody-coated Microplate Module (A)** to the mounting frame

Pipette 100 µL **Working Egg Standard** and **Working Sample Solution** into each well

- Incubate the wrapped microplate for 1 hour at 20-25°C (68-77°F)
- Wash the wells 6 times with **Washing Solution** \*

Dispense 100 µL **Enzyme-conjugated Antibody (C)** per well

- Incubate the microplate for 30 min at 20-25°C (68-77°F)
- Wash the wells 6 times with **Washing Solution** \*

Dispense 100 µL **Enzyme Substrate (D)** per well

- Incubate reaction for 30 min at 20-25°C (68-77°F) in the dark

Stop the enzyme reaction by adding 100 µL **Stop Solution (E)** per well

Measure absorbance at 450 nm

(Subtract the reading at a reference wavelength: any wavelength from 600 to 650 nm, inclusive)

Calculate the concentration of egg protein using the standard curve

\* The wells are emptied completely by inverting and tapping the microplate on a clean paper towel.



## RESULTS AND INTERPRETATION

1. Determine the mean absorbance for each set of **Working Sample Solution** and **Working Egg Standard**. Use graphing software, the built-in graphing feature of the microplate reader, or graph paper to construct a standard curve by plotting the mean absorbance value for each standard on the Y axis versus the corresponding **Working Egg Standard** concentration on the X axis.  
*Note: A standard curve should be generated for each assay.*
2. The egg protein concentration in the **Working Sample Solution** is interpolated from the standard curve using the mean absorbance value of each observation.
3. If the mean absorbance of **Working Sample Solution** is greater than the absorbance of the 50 ng/mL standard, increase the dilution of **Working Sample Solution** to 40-fold or greater as appropriate instead of 20-fold, and then assay again.

The egg protein content in a sample, in  $\mu\text{g/g}$ , can be estimated using the following formula:


$$\text{Egg protein content } (\mu\text{g/g}) = \text{OV} \times \text{Dilution A} \times \text{Dilution B} \times 1/1,000$$

OV: Observed value (ng/mL)

Dilution A: Dilution for ASSAY PROTOCOL a.2, nominal 20-fold.

Dilution B: Dilution for ASSAY PROTOCOL a.8, nominal 20-fold.

## TYPICAL STANDARD CURVE



## **STORAGE CONDITIONS AND EXPIRY OF KIT**

1. Store the kit at 2-8°C (35-46°F), but **DO NOT FREEZE!**
2. Use the kit at 20-25°C (68-77°F), while minimizing exposure time at this temperature. The kit must be returned to storage at 2-8°C (35-46°F) as soon as possible. Do not expose this kit to temperatures in excess of 25°C (77°F)!
3. Do not use the kit after the expiration date indicated on the outside box.

## **WARNINGS AND PRECAUTIONS**

### **(General)**

1. Egg protein is used in this kit. Users who are allergic to egg protein should handle kit reagents and perform the assay with great care. In case of an allergic reaction, such as sneezing or itching, seek medical attention if the reaction is serious or prolonged.
2. This product is recommended for use only by personnel trained in analytical sample handling techniques, and is meant to be used in accordance with good laboratory practices.
3. Direct contact with **Stop Solution (E)** should be strictly avoided. In case of contact, immediately flush affected area with copious amounts of water, and seek medical attention if necessary. In case of ingestion, call the poison control center and obtain first aid treatment.
4. Because all chemicals should be considered potentially hazardous, it is advisable to wear suitable protective clothing, goggles and gloves.
5. This product use 2-mercaptoethanol which is categorized as hazardous substance (CAS No.60-24-2), therefore perform the assay under the well-ventilated hood, and handle the reagents cautiously.
6. Because the assay is highly sensitive, all the procedures should be performed in a clean environment using uncontaminated equipment/devices and tubes/containers in order to minimize the risk of cross-contamination from previous analyses.

### **(ELISA)**

1. Do not combine reagents from different lots or other kits.
2. All reagents should be equilibrated at 20-25°C (68-77°F) before use.
3. A standard curve should be generated for each assay concurrently with the samples.
4. Assays should be performed at least in duplicate to ensure confidence in the measured values.
5. Care should be taken to pipette standard solutions and samples accurately.
6. Follow all procedures carefully.
7. Washing must be thorough in order to minimize background readings. Complete removal of reagents from the microplate wells is essential.
8. The enzyme substrate reaction should be performed shielded from light.

**Note:** *In the case of processed foods, the detection efficiency or sensitivity of the assay may decrease, due to denaturation and insolubilization of egg proteins. Therefore, food samples that give a negative result may still contain egg material which is either unreactive or present at concentrations below the limit of detection. It should not be assumed that such foods are egg free.*

## Reagent preparation

	Reagent	Preparation procedure
A	<b>Antibody-coated Microplate Module</b>	Equilibrate to 20-25°C (68-77°F) before opening Use immediately after opening
B	<b>Egg Standard</b> (50 ng/mL) <b>Working Egg Standards</b> (0.78, 1.56, 3.13, 6.25, 12.5 and 25 ng/mL)	Equilibrate to 20-25°C (68-77°F) Dilute <b>Egg Standard</b> (50 ng/mL) with <b>Diluent II</b> , sequentially
C	<b>Enzyme-conjugated Antibody</b>	Equilibrate to 20-25°C (68-77°F)
D	<b>Enzyme Substrate</b>	Equilibrate to 20-25°C (68-77°F) Avoid exposure to light
E	<b>Stop Solution</b>	Equilibrate to 20-25°C (68-77°F)
H	<b>Extraction Component A</b>	Equilibrate to 20-25°C (68-77°F), if necessary dissolve any remaining crystals in a water bath at 30-37°C (86-99°F)
	<b>Sample Extraction Solution</b>	Mix <b>Sample Buffer</b> , <b>Extraction Component A</b> , 2-Mercaptoethanol and distilled water at a ratio of 5:5:2:88
	<b>Diluent I</b>	Dilute <b>Sample Buffer</b> 20-fold with distilled water
	<b>Diluent II</b>	Dilute <b>Sample Extraction Solution</b> 20-fold with <b>Diluent I</b>
	<b>Washing Solution</b>	Dilute <b>Wash Buffer</b> 20-fold with distilled water

## TECHNICAL ASSISTANCE

For further technical assistance or troubleshooting advice, contact Morinaga Institute of Biological Science, Inc. or your local distributor.

Morinaga Institute of Biological Science, Inc.

Telephone: +81-45-791-7674

Fax: +81-45-791-7675

E-mail: info@miobs.com

## WARRANTY

Morinaga Institute of Biological Science, Inc. makes no warranty of any kind, either expressed or implied, except that the materials from which its products are made are of standard quality. Buyer assumes all risk and liability resulting from the use of this product. There is no warranty of merchantability of the product, or of the fitness of the product for any purpose. Morinaga Institute of Biological Science, Inc. agrees to replace any defective product, but expressly disclaims liability for damages, including special or consequential damages, or expenses arising directly or indirectly from the use of this product.